zeta hygiene



Cleaning and disinfection plan

WHICH		WHERE	WHEN	HOW	SPECTRUM OF ACTION
Zeta 1 Ultra	Account of the second of the s	Disinfection and cleaning of ordinary surgical instruments and burs	After possible contamination.	2% DILUTION (broad disinfection) Prepare the solution: dose 20 ml per litre of water. Disinfect the instruments: soak the instruments for 60 minutes in the static bath or for 30 minutes in the ultrasonic bath at 35°C. 1% DILUTION (rapid disinfection, with limited spectrum of action) Prepare the solution: dose 10 ml per litre of water. Disinfect the instruments: soak the instruments for 15 minutes in the static bath. At the end of the action time, remove the instruments from the dilution bath and rinse thoroughly.	2% DILUTION (BROAD DISINFECTION) STATIC BATH: Bactericidal: EN 13727, EN 14561 (S. aureus, P. aeruginosa, E. hirae) Yeasticidal: EN 13624, EN 14562 (C. albicans) Mycobactericidal, including tuberculocidal: EN 14348, EN 14563 (M. terrae, M. avium) Virucidal: EN 14476 (Poliovirus, Adenovirus, Norovirus, including HIV, HBV, HCV, Ebola, Herpes simplex and all human and animal influenza viruses) Bactericidal, yeasticidal (15'): VAH ULTRASONIC BATH: Bactericidal, yeasticidal, fungicidal, mycobactericidal incl. tuberculocidal, virucidal Tested under dirty conditions.
Zeta 2 Enzyme	March 19 19 19 19 19 19 19 19 19 19 19 19 19	Disinfection and cleaning of ordinary surgical instruments and burs	After possible contamination.	Prepare the solution: dose 20 g (2 measuring spoons) per litre of water. Wait 15 minutes. Disinfect the instruments: soak the instruments in the solution for 10 minutes. At the end of the action time, remove the instruments from the dilution bath and rinse thoroughly. The prepared solution remains stable for 8 hours.	Bactericidal: EN 13727, EN 14561 (S. aureus, P. aeruginosa, E. hirae) Yeasticidal: EN 13624, EN 14562 (C. albicans) Mycobactericidal, including tuberculocidal: EN 14348, EN 14563 (M. terrae, M. avium) Virucidal: EN 14476 (Poliovirus, Adenovirus, Norovirus, including HIV, HBV, HCV) Tested under dirty conditions.
Zeta 2 Sporex	Remark to a larger	Cold chemical sterilisation of ordinary surgical instruments and burs	After possible contamination, after disinfection.	Prepare the solution: dose 20 g (3 measuring spoons) per litre of water. Wait 15 minutes. Sterilize the instruments: soak the instruments in the solution for 10 minutes. Rinse thoroughly and dry. The prepared solution remains stable for 24 hours but should be renewed at the start of each working day.	Bactericidal: EN 13727, EN 14561 (S. aureus, P. aeruginosa, E. hirae) Fungicidal: EN 13624, EN 14562 (C. albicans, A. niger) Mycobactericidal, including tuberculocidal: EN 14348, EN 14563 (M. terrae, M. avium) Virucidal: EN 14476 (Poliovirus, Adenovirus, Norovirus, Parvovirus, including HIV, HBV, HCV) Sporicidal: EN 13704 (B. subtilis) Tested under dirty conditions.
Zeta 3 Soft	The state of the s	Disinfection and cleaning of surfaces of medical devices	After possible contamination. As required.	Spray the product over the surface to be disinfected directly or using a wipe*. Thoroughly rub the entire surface to be disinfected before the product evaporates completely. Leave to dry. *Before applying, check the compatibility of the alcohol-based product on a hidden part of the surface to be treated. Do not treat delicate or alcohol-sensitive surfaces.	Bactericidal: EN 13727, EN 14561 (S. aureus, P. aeruginosa, E. hirae) Yeasticidal: EN 13624, EN 14562 (C. albicans) Tuberculocidal: EN 14348, EN 14563 (M. terrae) Virucidal: EN 14476 (Poliovirus, Adenovirus, Norovirus, Parvovirus, including HIV, HBV, HCV) Tested under dirty conditions.
Zeta 3 Wipes TOTAL	Marie Constitution of the	Disinfection and cleaning of surfaces of medical devices	After possible contamination. As required.	Thoroughly rub the surface with the wipe ensuring the complete coverage of the surface to be disinfected*. Leave to dry. *Before applying, check the compatibility of the alcohol-based product on a hidden part of the surface to be treated. Do not treat delicate or alcohol-sensitive surfaces.	Bactericidal: EN 13727, EN 14561 (S. aureus, P. aeruginosa, E. hirae) Yeasticidal: EN 13624, EN 14562 (C. albicans) Tuberculocidal: EN 14348, EN 14563 (M. terrae) Virucidal: EN 14476 (Poliovirus, Adenovirus, Norovirus, Parvovirus, including HIV, HBV, HCV) Tested under dirty conditions. Standards apply to the solution. Effectiveness of wipes (According to the prEN 16615 method, 10' contact): R>4Log vs. S. aureus, E. hirae R>3Log vs. C. albicans, P. aeruginosa
Zeta 3 Foam		Disinfection and cleaning of delicate surfaces of medical devices	After possible contamination. As required.	Spray the product over the surface to be disinfected directly or using a wipe*. Distribute evenly with a wipe to ensure complete coverage of the surfaces to be disinfected. Leave to dry. *Before applying, check the compatibility of the product on a hidden part of the surface to be treated.	Bactericidal: EN 13727, EN 14561 (S. aureus, P. aeruginosa, E. hirae) Yeasticidal: EN 13624, EN 14562 (C. albicans) Tuberculocidal: EN 14348, EN 14563 (M. terrae) Virucidal: EN 14476 (Poliovirus, Adenovirus, Norovirus, Parvovirus, including HIV, HBV, HCV) Tested under dirty conditions.
Zeta 3 Wipes POP-UP	Page 1 Appen 100	Disinfection and cleaning of delicate surfaces of non-invasive medical devices	After possible contamination. As required.	Thoroughly rub the surface with the wipe ensuring the complete coverage of the surface to be disinfected*. *Before applying, check the compatibility of the product on a hidden part of the surface to be treated.	Bactericidal: EN 13697, EN 1276*, EN 14561*, EN 14561 (MRSA) Fungicidal: EN 13697 (A. niger), EN 14562 (A. fumigatus) Yaesticidal: EN 1650*, EN 13624, EN 13697, EN 14562 (C. albicans) Tuberculocidal: EN 14348, EN 14563 (M. terrae) Virucidal: EN 14476 (HBV, HCV, Adenovirus, Coronavirus, Norovirus, VRS, Polyomavirus, H1N1, Rotavirus), EN 14476* (HSV) Tested under clean conditions. *Tested under dirty conditions.
Zeta 4 Wash	The state of the s	Cleaning of washable surfaces	At the end of the working day.	Prepare the solution: dilute 1-2 dosage caps per litre of water. Wash the floor as usual. For persistent dirt, apply a few drops of the product on a damp cloth and rub before washing.	Cleaner
Zeta 5 Power Act	Martine or Commerce of Commerc	Disinfection and cleaning of suction units	At least at the end of the working day and after each invasive surgical intervention.	Rinse the suction system by drawing up 2 litres of cold water. Prepare the solution: fill a bucket with water and dose 10 ml per litre of water. Then mix the solution obtained. Disinfect the suction unit: draw the disinfectant solution, alternating it with air in the suction tubes. For a rapid disinfection, let it work for 15 minutes (excluding tuberculocidal action). For a complete disinfection, let it work overnight. Pour the remaining disinfectant solution (at least 250 - 300 ml) into the spittoon. Rinse the suction system again by drawing up 2 litres of cold water.	Bactericidal: EN13727, EN 14561 (S. aureus, P. aeruginosa, E. hirae) Yaesticidal: EN 13624, EN 14562 (C. albicans) Tuberculocidal: EN 14348 (M. terrae) Limited virucidal: prEN 16777, DVV/RKI vs. encapsulated and lipophilic viruses including blood viruses (HIV, HBV, HCV), Herpes simplex and families of virus such as the Orthomyxoviridae (including all human and animal influenza viruses such as H5N1 and H1N1), Filoviridae (Ebola virus) and Paramyoviridae (measles virus) Bactericidal, yeasticidal: VAH Tested under dirty conditions.
Zeta 7 Solution	Mary Construction of the C	Disinfection of impressions	Immediately after taking the impression before transporting it to the lab. After receipt from the lab.	Rinse the impression thoroughly under running water. Prepare the solution: dose 10 ml per litre of water. Disinfect the impression: soak in the solution for 10 minutes. At the end of the action time, remove the impression from the dilution bath and rinse thoroughly.	Bactericidal: EN 13727 (S. aureus, P. aeruginosa, E. hirae) Yeasticidal: EN 13624 (C. albicans) Tuberculocidal: EN 14348, EN 14563 (M. terrae) Virucidal against enveloped viruses: EN 14476 (Vaccinia Virus including HIV, HBV, HCV) Tested under dirty conditions.
Zeta 7 Spray	American Control of the Control of t	Disinfection of impressions	Immediately after taking the impression before transporting it to the lab. After receipt from the lab.	Rinse the impression thoroughly under running water. Disinfect the impression: spray all over the impression (both sides). Let the product evaporate for 3 minutes.	Bactericidal: EN 13727 (S. aureus, P. aeruginosa, E. hirae) Yeasticidal: EN 13624 (C. albicans) Tuberculocidal: EN 14348, EN 14563 (M. terrae) Virucidal: EN 14476 (Poliovirus, Adenovirus, Norovirus, Parvovirus, including HIV, HBV, HCV) Tested under dirty conditions.
Zeta 6 Hydra	Account of the control of the contro	Hand hygiene	Before and after every treatment. After every contamination. Additionally as required.	Wet your hands and pour the product on the palm of one hand, then rub the product thoroughly over your hands before rinsing and drying.	Liquid cleaner for frequent hand washing, containing nutrient (Hydrolised Rice Protein), protective (Glyceril Oleate) and hydrating substances (Glycerin). Specially formulated for delicate skins.





DISINFECTION



Zeta 2 Enzyme Concentrated 2²⁰







COLD CHEMICAL STERILIZATION

Zeta 2 Sporex Concentrated 2²⁰







CLEANING









SURFACES







SURFACES OF MEDICAL DEVICES

Zeta 3 Wipes TOTAL



DELICATE SURFACES



Zeta 3 Wipes POP-UP



WASHABLE SURFACES









SUCTION UNITS

Zeta 5 Power Act Concentrated 1

























IMPRESSIONS

Zeta 7 Solution Concentrated 120













Zeta 7 Spray







Spray all over the impression (both sides)





HANDS

Zeta 6 Hydra





















